

Specification

Acetic Acid 99.7%

ACS Grade

Property

Assay
Color
Dilution Test
Acetic Anhydride
Chloride
Residue After Evaporation
Sulfate
Titrable Base
Reducing Substances Dichromate
Reducing Substances Permanganate
Heavy Metals
Iron

Specification

99.7% minimum
10 APHA maximum
To Pass Test
0.01% maximum
1 ppm maximum
10 ppm maximum
1 ppm maximum
0.0004 meq/g maximum
To Pass Test
To Pass Test
0.5 ppm maximum
0.2 ppm maximum