



## Certificate of Analysis

### Acetic Acid 0.1N Solution

#### PharmaPUR™

#### GMP Manufactured Product

<u>Property</u>	<u>Specifications</u>	<u>Analysis</u>	
Assay Normality	0.095-0.105 N	0.102	N
Appearance	Clear, colorless liquid; free from suspended matter	Pass	
Sulfate	Passes Test	Pass	
Chloride	25 ppm maximum	<25	ppm
Non Volatile Residue	1 mg/20ml maximum	<1	mg
Heavy Metals	5 ppm maximum	<5	ppm
Iron	5 ppm maximum	<5.000	ppm
Identification	Passes Test	Pass	

Made from Purified Water, USP and Glacial Acetic Acid, USP; which also meets EP and JP specifications.

Lot Number 24991131501

By Stephanie Pinho  
Stephanie Pinho  
Laboratory Manager

Manufacture Date 04/24/2015  
Expiration Date 04/30/2017

Part Number 249992GMP