



Certificate of Analysis

Hydrochloric Acid 3.00N Solution

PharmaPUR™

GMP Manufactured Product

<u>Property</u>	<u>Specifications</u>	<u>Analysis</u>	
Assay Normality	2.95-3.05 N	2.96	N
Appearance	Clear, colorless liquid	Pass	
Sulfate	Passes NF test	Pass	
Residue after Ignition	2 mg maximum	<2.00	mg
Free Bromine or Chlorine	Passes NF test	Pass	
Sulfite	Passes NF test	Pass	
Heavy Metals	5 ppm maximum	<5	ppm
Identification	Passes NF test	Pass	

Made from USP Purified Water and Hydrochloric Acid, NF

Lot Number 63742851501

By Stephanie Pinho
Stephanie Pinho
Laboratory Manager

Manufacture Date 10/15/2015
Expiration Date 10/31/2017