



# PURITAN PRODUCTS

## Certificate of Analysis

### Hydrochloric Acid 0.24 N Solution

#### PharmaPUR™

<u>Property</u>	<u>Specifications</u>	<u>Analysis</u>	
Assay Normality	0.23-0.25N	0.240	N
Sulfate	Passes Test	Pass	ppm
Residue after Ignition	80 ppm maximum	<80.00	ppm
Bromide or Iodide	Passes Test	Pass	
Free Bromine or Chlorine	Passes Test	Pass	
Sulfite	Passes Test	Pass	
Heavy Metals	5 ppm maximum	<5	ppm
Identification	Passes Test	Pass	

Lot Number 63791321501

By Stephanie Pinho  
Stephanie Pinho  
Laboratory Manager

Manufacture Date 05/12/2015  
Expiration Date 05/31/2017

Part Number 637955GMP