



PURITAN PRODUCTS

Certificate of Analysis

Sulfuric Acid 5.0 N Solution

PharmaPUR™

GMP Manufactured Product

<u>Property</u>	<u>Specifications</u>	<u>Analysis</u>	
Appearance of Solution	Clear and Colorless	Pass	
Assay Normality	4.9 - 5.1 N	4.980	N
Identification A	To Pass Test	Pass	
Identification B	To Pass Test	Pass	
Limit of Nitrate	To Pass Test	Pass	
Residue after Ignition	20 ppm maximum	<20.00	ppm
Chloride	20 ppm maximum	<20	ppm
Reducing Substances	To Pass Test	Pass	
Arsenic	0.4 ppm maximum	<0.400	ppm
Heavy Metals	2 ppm maximum	<2	ppm
Iron	10 ppm maximum	<10.000	ppm

Made from USP Purified water and Sulfuric Acid, NF. This product meets the USP guideline on residual solvents.

Lot Number 83861311501

By Stephanie Pinho

Stephanie Pinho
Laboratory Manager

Manufacture Date 05/11/2015
Expiration Date 05/31/2018

Part Number 838615