



## Certificate of Analysis

### Buffer Standard Solution pH 7.00 Yellow

**PharmaPUR™**  
**NIST TRACEABLE**

<u>Property</u>	<u>Specifications</u>	<u>Analysis</u>
Appearance	Clear, Yellow Solution	Pass
pH at 25 deg C	7.00 +/- 0.01	7.00

Store at room or refrigerated temperatures

Our results are reported as described in Guide to Expression of Uncertainty in Measurement, ISO, Geneva, 1993, with a coverage factor of 2 (confidence level of 95%). The expanded uncertainty is  $\pm 0.02$  pH for products in the pH range 0.50 through 10.00. Above 10.00 pH, the expanded uncertainty is  $\pm 0.05$  pH.

Lot Number 97470171701

By   
Amy Yencho

Laboratory Manager

Manufacture Date 01/17/2017  
Expiration Date 01/31/2019