



## Certificate of Analysis

### Buffer Standard Solution pH 7.00 Yellow

**PharmaPUR™**  
**NIST TRACEABLE**

<u>Property</u>	<u>Specifications</u>	<u>Analysis</u>
Appearance	Clear, Yellow Solution	Pass
pH at 25 deg C	7.00 +/- 0.01	7.01

Store at room or refrigerated temperatures

Our results are reported as described in Guide to Expression of Uncertainty in Measurement, ISO, Geneva, 1993, with a coverage factor of 2 (confidence level of 95%). The expanded uncertainty is  $\pm 0.02\text{pH}$  for products in the pH range 0.50 through 10.00. Above 10.00pH, the expanded uncertainty is  $\pm 0.05\text{pH}$ .

Lot Number 97470841901

By

John Sinclair  
Site Quality Manager

Manufacture Date 03/25/2019  
Expiration Date 03/31/2021