



## Certificate of Analysis

### Buffer pH 7.0 Yellow

#### NIST TRACEABLE

| <u>Property</u> | <u>Specifications</u>  | <u>Analysis</u> |
|-----------------|------------------------|-----------------|
| Appearance      | Clear, Yellow Solution | Pass            |
| pH at 25 deg C  | 7.00 +/- 0.01          | 7.00            |

Store at room or refrigerated temperatures

Our results are reported as described in Guide to Expression of Uncertainty in Measurement, ISO, Geneva, 1993, with a coverage factor of 2 (confidence level of 95%). The expanded uncertainty is  $\pm 0.02\text{pH}$  for products in the pH range 0.50 through 10.00. Above 10.00pH, the expanded uncertainty is

Lot Number 97472791401

By Stephanie Pinho Manufacture Date 10/06/2014  
Stephanie Pinho  
Laboratory Manager

Part Number 9747GMP