



## Certificate of Analysis

### Buffer Standard Solution pH 10.00 Blue

#### PharmaPUR™ NIST TRACEABLE

| <u>Property</u> | <u>Specifications</u> | <u>Analysis</u> |
|-----------------|-----------------------|-----------------|
| Appearance      | Clear, Blue Solution  | Pass            |
| pH at 25 deg C  | 10.00 +/- 0.01        | 10.01           |

Store at room or refrigerated temperatures

Our results are reported as described in Guide to Expression of Uncertainty in Measurement, ISO, Geneva, 1993, with a coverage factor of 2 (confidence level of 95%). The expanded uncertainty is  $\pm 0.02$  pH for products in the pH range 0.50 through 10.00. Above 10.00 pH, the expanded uncertainty is  $\pm 0.05$  pH.

Lot Number 97480391701

By   
Amy Yencho

Laboratory Manager

Manufacture Date 02/09/2017  
Expiration Date 02/28/2018