



Certificate of Analysis

Buffer Standard Solution pH 10.00 Blue

PharmaPUR™ NIST TRACEABLE

<u>Property</u>	<u>Specifications</u>	<u>Analysis</u>
Appearance	Clear, Blue Solution	Pass
pH at 25 deg C	10.00 +/- 0.01	10.01

Store at room or refrigerated temperatures

Our results are reported as described in Guide to Expression of Uncertainty in Measurement, ISO, Geneva, 1993, with a coverage factor of 2 (confidence level of 95%). The expanded uncertainty is ± 0.02 pH for products in the pH range 0.50 through 10.00. Above 10.00 pH, the expanded uncertainty is ± 0.05 pH.

Lot Number 97480711801

By 
Amy Yencho

Laboratory Manager

Manufacture Date 03/12/2018
Expiration Date 03/31/2019