



Certificate of Analysis

Buffer Standard Solution pH 10.00 Blue

PharmaPUR™
NIST TRACEABLE

<u>Property</u>	<u>Specifications</u>	<u>Analysis</u>
Appearance	Clear, Blue Solution	Pass
pH at 25 deg C	10.00 +/- 0.01	10.00

Store at room or refrigerated temperatures

Our results are reported as described in Guide to Expression of Uncertainty in Measurement, ISO, Geneva, 1993, with a coverage factor of 2 (confidence level of 95%). The expanded uncertainty is ± 0.02 pH for products in the pH range 0.50 through 10.00. Above 10.00 pH, the

Lot Number 97482941601

By Stephanie Pinho
Stephanie Pinho

Manufacture Date 10/24/2016
Expiration Date 10/31/2017

Laboratory Manager