



Certificate of Analysis

Buffer Standard Solution pH 4.00 Red

PharmaPUR™
NIST TRACEABLE

Property

pH at 25 deg C

Specifications

4.00 +/- 0.01

Analysis

4.01

Store at room or refrigerated temperatures

Our results are reported as described in Guide to Expression of Uncertainty in Measurement, ISO, Geneva, 1993, with a coverage factor of 2 (confidence level of 95%). The expanded uncertainty is $\pm 0.02\text{pH}$ for products in the pH range 0.50 through 10.00. Above 10.00pH, the expanded uncertainty is $\pm 0.05\text{pH}$.

Lot Number 97490571901

By

John Sinclair
Site Quality Manager

Manufacture Date 02/26/2019
Expiration Date 02/28/2021