



# PURITAN PRODUCTS

## Certificate of Analysis

### Buffer pH 4.00 Red

#### NIST TRACEABLE

<u>Property</u>	<u>Specifications</u>	<u>Analysis</u>
Appearance	Clear, Red Solution	Pass
pH at 25 deg C	4.00 +/- 0.01	4.00

Store at room or refrigerated temperatures

Our results are reported as described in Guide to Expression of Uncertainty in Measurement, ISO, Geneva, 1993, with a coverage factor of 2 (confidence level of 95%). The expanded uncertainty is  $\pm 0.02\text{pH}$  for products in the pH range 0.50 through 10.00. Above 10.00pH, the

Lot Number 97492821401

By Stephanie Pinho

Stephanie Pinho  
Laboratory Manager

Manufacture Date 10/09/2014

Expiration Date 10/31/2016

Part Number 974992GMP