



Certificate of Analysis

Buffer Standard Solution pH 4.00 Red

PharmaPUR™
NIST TRACEABLE

<u>Property</u>	<u>Specifications</u>	<u>Analysis</u>
Appearance	Clear, Red Solution	Pass
pH at 25 deg C	4.00 +/- 0.01	4.00

Store at room or refrigerated temperatures

Our results are reported as described in Guide to Expression of Uncertainty in Measurement, ISO, Geneva, 1993, with a coverage factor of 2 (confidence level of 95%). The expanded uncertainty is $\pm 0.02\text{pH}$ for products in the pH range 0.50 through 10.00. Above 10.00pH, the

Lot Number 97493071601

By <u>Stephanie Pinho</u>	Manufacture Date	11/02/2016
Stephanie Pinho	Expiration Date	11/30/2018
Laboratory Manager		